



INNOVATIVE SOLUTIONS FOR CRITICAL ENVIRONMENTS

Valutek, Inc

220 South 9th St.

Phoenix, Arizona, 85034, United States

P +1 602.256.0540 | F +1 602.252.1972

valutek.com | quality@valutek.com

CERTIFICATE OF ANALYSIS

Part Number:	VTPNWLS-1818
Product Type:	Polyester Wiper
Edge Treatment:	Laser Sealed
Lot Number:	240930-101002
Sample Received Date:	10-07-2024
Manufacturing Date:	09-30-2024
Test Date:	10-15-2024

Page 1 of 2

Physical Properties

Test	Method	Specification	Actual	Result
Dimensions	Tape Measure	18" x 18" ($\pm 0.5"$)	18.0 x 18.0	Pass
Basic Weight	TAPPI T-410	135 $\pm$ 5 g/m ²	137	Pass

Absorbency

Test	Method	Specification	Actual	Result
Sorptive Efficiency	IENT-RP.CC004.3	Min. 2 ml/g	2.2	Pass
Sorptive Capacity	IENT-RP.CC004.3	Min. 310 ml/m ²	316	Pass
Sorptive Rate	IENT-RP.CC004.3	< 1 sec	0.5	Pass

Cleanliness

Test	Method	Specification	Actual	Result
LPC ($\geq 0.5\mu\text{m}$)	IENT-RP.CC004.3	< 10 x 10 ⁶ particles/m ²	8.35	Pass
TNVR - IPA	IENT-RP.CC004.3	Max. 0.1 g/m ²	0.010	Pass
TNVR – DI H2O	IENT-RP.CC004.3	Max. 0.02 g/m ²	0.008	Pass

Extractables

Test Method: IEST-RP.CC004.3

Ions	Specification	Actual	Result
Sodium	Max. 0.2 ppm	ND	Pass
Potassium	Max. 0.2 ppm	ND	Pass
Calcium	Max. 0.5 ppm	ND	Pass
Magnesium	Max. 0.2 ppm	0.019	Pass
Ammonium	Max. 0.3 ppm	0.052	Pass
Nitrate	Max. 0.3 ppm	ND	Pass
Fluoride	Max. 0.2 ppm	ND	Pass
Nitrite	Max. 0.2 ppm	ND	Pass
Bromide	Max. 0.2 ppm	ND	Pass
Phosphate	Max. 0.3 ppm	0.041	Pass
Chloride	Max. 0.2 ppm	0.021	Pass
Sulphate	Max. 0.2 ppm	0.045	Pass

Fourier Transform Infra-Red Test

FTIR	Method	Specification	Actual	Result
Silicone, Amide & DOP	IEST-RP.CC004-3	Absent	Absent	Pass

Notes:



Tested by:
Renugaa Krishnan
 Laboratory Coordinator



Approved by:
Vijendraan Venkitasalam
 Quality Manager